

OK to Enter  
21 July 2005

**IN THE CLAIMS:**

Claim 1 has been amended as follows:

1. (Currently Amended) A device to install and remove a structural component of a medical installation, said medical installation having a height-adjustable patient bed separate from said structural component, said device comprising:

a two-part guide system attachable to said patient bed and to said structural component;

a first of the two parts of said guide system comprising a guide rail and a second of said two parts of said guide system comprising a guide groove, ~~said guide~~ patient bed rail being adapted to receive said patient bed guide rail thereon; and

said guide system, upon temporary, detachable placement of said structural component on said guide system on said patient bed, guiding said structural component by sliding along said guide rail relative to said medical installation.

2. (Original) A device as claimed in claim 1 wherein said guide groove is in said structural component.

3. (Original) device as claimed in claim 1 wherein said structural component has a bearing support attached thereto, and wherein said guide groove is in said bearing support.

4. (Original) A device as claimed in claim 3 wherein said bearing support is comprised of plastic.

5. (Original) A device as claimed in claim 1 wherein said guide system comprises an attachment element for attaching said guide rail to said patient bed.

OK to Enter  
Jul 21 July 2013

6. (Original) A device as claimed in claim 1 wherein said guide rail is a first guide rail, and wherein said guide system comprises a second guide rail mounted on said medical device that, with appropriate positioning of said patient bed, forms an extension of said first guide rail.

7. (Original) A device as claimed in claim 6 wherein said second guide rail is comprised of plastic.

8. (Original) A device as claimed in claim 1 wherein said medical device is a magnetic resonance tomography device.

9. (Original) A device as claimed in claim 8 wherein said structural component is a radio-frequency body antenna of said magnetic resonance tomography device.

10. (Original) A device as claimed in claim 1 wherein said structural component is a gradient system of said magnetic resonance tomography device.

11. (Original) A device as claimed in claim 10 wherein said guide rail is a first guide rail, and wherein said guide system comprises a second guide rail mounted to said gradient system, said second guide rail, given appropriate positioning of said patient bed, forming an extension of said first guide rail.

12. (Previously Presented) A method for installing and removing a structural component of a medical device comprising the steps of:

positioning a first part of a guide system at a height-adjustable patient bed of the medical device;

forming a second part of the guide system on a structural component, separate from said bed and temporarily detachably engaging said first part of said guide system with said second part of said guide system

OK to Enter  
WU 21 July 2005

with said structural component on said guide system on said patient bed; and

adjusting the height of the patient bed to selectively raise and lower the structural component therein relative to said medical device, and sliding said structural component along said first and second parts of said guide system to install or remove said component relative to said medical device.

13. (Previously Presented) A magnetic resonance tomography device comprising:

a magnetic resonance scanner having a radio-frequency body antenna and a gradient system;

a height-adjustable patient bed adapted to receive a patient thereon to move said patient into and out of said magnetic resonance scanner; and

a device for installing and removing a structural component, selected from the group consisting of said radio-frequency body antenna and said gradient system, relative to said magnetic resonance scanner, said device comprising a two-part guide system having a first part attached to said height adjustable patient bed and a second part attached to said structural component, said first part comprising a guide rail and said second part comprising a guide groove temporarily detachably engageable with said guide rail allowing said structural component, when placed on said patient bed, to be slid along said guide rail relative to said magnetic resonance scanner.